UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/531,818	06/01/2006	Ghaleb Slater	PA1906 US2	3928	
/-	7590 05/15/2008 VASCULAR, INC.	;	EXAMINER		
IP LEGAL DEI	PARTMENT		PREBILIC, PAUL B		
3576 UNOCAL SANTA ROSA	=		ART UNIT	PAPER NUMBER	
			3774		
			NOTIFICATION DATE	DELIVERY MODE	
			05/15/2008	ELECTRONIC	

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

rs.vasciplegal@medtronic.com

Office Action Communication		Application No.		Applicant(s)					
		10/531,818	3	SLATER, GHALEB					
Office Action Summary			Examiner		Art Unit				
			Paul B. Pre	bilic	3774				
Period fo	The MAILING DATE of this commur or Reply	nication appe	ears on the	cover sheet with the o	correspondence ad	ddress			
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).									
Status									
1) 又	Responsive to communication(s) file	ed on <i>14 Fe</i>	hruary 200	8					
•	Responsive to communication(s) filed on <u>14 February 2008</u> . This action is FINAL . 2b) This action is non-final.								
3)		<i>,</i> —			osecution as to the	e merits is			
٥,١	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.								
Dispositi	on of Claims								
4)⊠	Claim(s) 8-16 is/are pending in the	application.							
•	4a) Of the above claim(s) is/are withdrawn from consideration.								
· · · · · · · · · · · · · · · · · · ·	5) Claim(s) is/are allowed.								
· · · · ·	Claim(s) <u>10-14 and 16</u> is/are rejected.								
•	7)⊠ Claim(s) <u>8,9 and 15</u> is/are objected to. 3)□ Claim(s) are subject to restriction and/or election requirement.								
ا ا	Claim(s) are subject to restin	Ction and/or	election re	quirement.					
Applicati	on Papers								
9)	The specification is objected to by th	ne Examiner	r.						
10)	The drawing(s) filed on is/are	: a) <u>□</u> acce	epted or b)[objected to by the	Examiner.				
	Applicant may not request that any object	ection to the d	drawing(s) be	held in abeyance. Se	e 37 CFR 1.85(a).				
	Replacement drawing sheet(s) including	g the correction	on is require	d if the drawing(s) is ob	jected to. See 37 C	FR 1.121(d).			
11)	11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.								
Priority ι	ınder 35 U.S.C. § 119								
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 									
2) Notic 3) Inform	t(s) e of References Cited (PTO-892) e of Draftsperson's Patent Drawing Review (I nation Disclosure Statement(s) (PTO/SB/08) r No(s)/Mail Date			4) Interview Summary Paper No(s)/Mail D 5) Notice of Informal F 6) Other:	ate				

Application/Control Number: 10/531,818 Page 2

Art Unit: 3774

Claim Objections

Claims 8 and 9 are objected to because of the following informalities. On the second to last and last lines of claims 8 and 9, the terminology "treatment device" lacks clear antecedent basis. The Examiner suggests changing this language to ---device for treating cardiac valve regurgitation--- in order to overcome this objection. Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 10-14 and 16 are rejected under 35 U.S.C. 102(e) as being anticipated by Mathis et al (US 2003/0212453). Mathis anticipates the claim language where:

- (a) the compression member as claimed is the anchor member (34) of Mathis (see Figure 9 and paragraph [0013],
 - (b) the tube as claimed is crimp (70) (see Figures 5, 11, and 12),
 - (c) the sleeve as claimed is catheter (50) (see Figure 3),
 - (d) the positioning step is best shown in Figures 1 and 2,
- (e) the relative rotation step is about the arc formed around the heart valve (20) by coronary sinus (14) (see Figure 2), and

Application/Control Number: 10/531,818 Page 3

Art Unit: 3774

(f) the side port as claimed is the hole on the left side or end of catheter (50).

With regard to claim 13, the catheter (50) of Mathis is the sleeve that is a delivery catheter as claimed.

Allowable Subject Matter

Claim 16 is objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Claims 8 and 9 would be allowable if the objectionable language were corrected.

Conclusion

Applicant should specifically point out the support for any amendments made to the disclosure, including the claims (MPEP 714.02 and 2163.06). Due to the procedure outlined in MPEP 2163.06 for interpreting claims, it is noted that other art may be applicable under 35 USC 102 of 35 USC 103(a) once the aforementioned issue(s) is/are addressed.

Applicant is respectfully requested to provide a list of all copending applications that set forth similar subject matter to the present claims. A copy of such copending claims is respectfully requested in response to this Office action if the application is not stored in image format (i.e. the IFW system) or published.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Examiner Paul B. Prebilic whose telephone number is (571) 272-4758. He can normally be reached on 6:30-5:00 M-Th.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on 571-272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should

Application/Control Number: 10/531,818

Art Unit: 3774

you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/Paul Prebilic/ Paul Prebilic Primary Examiner Art Unit 3774 Page 4